

Calibration Certificate of Mass

Calibration Date: August 22, 2022

Certificate Number: 2022-114-1

Submitted By: FSCP Area 20
3721 West Cuming St.
Lincoln, NE 68524

Point of Contact: Kurt Wenninghoff
Ph. 402-471-3422
email: kurt.wenninghoff@nebraska.gov
PO Number: N/A

Test Item(s): Cast weights	Artifact(s) Description:	Date Received: August 19, 2022
ID / Asset Number: Area 20		Serial Number(s): See Next Page
Manufacture: Various		Class Specification: NIST Class F
Material: Cast Iron		Condition: Fair (significant wear)

Reference Standards Used: NSL lb standards	Procedure Used: NIST HB 6969, SOP 8 (2019) Metrologist: JPL	Equipment Used: Mettler XPR32003 Mettler XP 604
--	--	--

Environmental Cond. **Temp:** 22.9 °C **Pressure:** 733 mmHg **Relative Humidity:** 47.3 %

Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and the uncertainty exceed 95% of the maximum permissible error. All of the tolerances and design specifications (except density, hardness and magnetism) were evaluated according to ASTM E617 (2018) and/or NIST HB 105-1 (2019) for the artifacts designated class.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm³ at 20 °C.
- It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2022), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales.

Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: August 22, 2022

Certificate Number: 2022-114-1


Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm ³)
25 lb	WM25-1	0.72	N	0.72	0.14	2	1.1	7.2
25 lb	WM25-2	0.38	N	0.38	0.14	2	1.1	7.2
25 lb	WM25-3	0.28	N	0.28	0.14	2	1.1	7.2
25 lb	WM25-4	1.23	Y	-0.01	0.14	2	1.1	7.2
25 lb	WM25-5	-0.01	N	-0.01	0.14	2	1.1	7.2
25 lb	WM25-6	1.06	Y	0.45	0.14	2	1.1	7.2
25 lb	WM25-7	1.09	Y	0.41	0.14	2	1.1	7.2
25 lb	WM25-8	-0.06	N	-0.06	0.14	2	1.1	7.2
25 lb	WM25-9	1.06	Y	0.74	0.14	2	1.1	7.2
25 lb	WM25-10	0.50	N	0.50	0.14	2	1.1	7.2
25 lb	WM25-11	0.96	N	0.96	0.14	2	1.1	7.2
25 lb	WM25-12	1.20	Y	0.36	0.14	2	1.1	7.2
25 lb	WM25-13	0.54	N	0.54	0.14	2	1.1	7.2
25 lb	WM25-14	0.18	N	0.18	0.14	2	1.1	7.2
25 lb	WM25-15	0.03	N	0.03	0.14	2	1.1	7.2
25 lb	WM25-16	-0.68	N	-0.68	0.14	2	1.1	7.2
25 lb	WM25-17	1.28	Y	0.55	0.14	2	1.1	7.2
25 lb	WM25-18	0.36	N	0.36	0.14	2	1.1	7.2
25 lb	WM25-19	0.47	N	0.47	0.14	2	1.1	7.2
25 lb	WM25-20	0.87	N	0.87	0.14	2	1.1	7.2
50 lb	A5C-13	-0.90	N	-0.90	0.28	2	2.3	7.2
50 lb	A5C-20	-2.52	Y	-1.23	0.28	2	2.3	7.2
50 lb	OPI-C67	-2.45	Y	-0.64	0.28	2	2.3	7.2
50 lb	WM-OPI-C85	-1.48	N	-1.48	0.28	2	2.3	7.2
1000 lb	A-1	13.5	N	13.5	5.8	2.019	45	7.2
1000 lb	A-3	0.1	N	0.1	5.8	2.019	45	7.2
1000 lb	A-4	-12.6	N	-12.6	5.8	2.019	45	7.2
1000 lb	A-7	-17.4	N	-17.4	5.8	2.019	45	7.2
1000 lb	A-8	-29.5	N	-29.5	5.8	2.019	45	7.2
1000 lb	A-9	-72.0	Y	5.7	5.8	2.019	45	7.2
1000 lb	A-10	-21.0	N	-21.0	5.8	2.019	45	7.2
1000 lb	A-14	-17.5	N	-17.5	5.8	2.019	45	7.2
1000 lb	A-17	-15.8	N	-15.8	5.8	2.019	45	7.2
1000 lb	A-18	-11.2	N	-11.2	5.8	2.019	45	7.2
1000 lb	A-20	-38.6	N	-38.6	5.8	2.019	45	7.2
1000 lb	2189	-34.1	N	-34.1	5.8	2.019	45	7.2
1000 lb	2190	-4.0	N	-4.0	5.8	2.019	45	7.2
1000 lb	2191	-33.5	N	-33.5	5.8	2.019	45	7.2
1000 lb	2192	16.1	N	16.1	5.8	2.019	45	7.2
1000 lb	2193	9.7	N	9.7	5.8	2.019	45	7.2
1000 lb	2194	-13.2	N	-13.2	5.8	2.019	45	7.2
1000 lb	2195	-28.2	N	-28.2	5.8	2.019	45	7.2
1000 lb	2196	6.9	N	6.9	5.8	2.019	45	7.2
1000 lb	2197	-18.7	N	-18.7	5.8	2.019	45	7.2

Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

e-signature is copy only

8/25/2022

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

Calibration Certificate of Mass

Calibration Date: August 23, 2022

Certificate Number: 2022-114-2

Submitted By: FSCP Area 20
3721 West Cuming St.
Lincoln, NE 68524

Point of Contact: Kurt Wenninghoff
Ph. 402-471-3422
email: kurt.wenninghoff@nebraska.gov
PO Number: N/A

Test Item(s): lb weight kit
Serial Number(s): WM-2D86
Manufacture: Rice Lake
Material: Stainless Steel

Artifact(s) Description:

Date Received: August 19, 2022
ID / Asset Number: Area 20
Class Specification: NIST Class F
Condition: Fair (significant wear)

Reference Standards Used:

NSL lb standards

Procedure Used:

NIST HB 6969, SOP 8 (2019)

Metrologist:

JPL

Equipment Used:

Sartorius CC10000S Mettler XPR 205
Sartorius CC 1201 Sartorius CCE6

Environmental Cond. **Temp:** 21.29 °C **Pressure:** 729.6 mmHg **Relative Humidity:** 51.07 %

Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and the uncertainty exceed 95% of the maximum permissible error. All of the tolerances and design specifications (except density, hardness and magnetism) were evaluated according to ASTM E617 (2018) and/or NIST HB 105-1 (2019) for the artifacts designated class.

- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm³ at 20 °C.
- It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2020), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales.

Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: August 23, 2022

Certificate Number: 2022-114-2

Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm ³)
5 lb	1	-0.082	n	-0.082	0.028	2	0.23	7.84
5 lb	2	-0.110	n	-0.110	0.028	2	0.23	7.84
5 lb	3	-0.103	n	-0.103	0.028	2	0.23	7.84
5 lb	4	-0.043	n	-0.043	0.028	2	0.23	7.84
5 lb	5	-0.061	n	-0.061	0.028	2	0.23	7.84
1 lb	1	0.0077	n	0.0077	0.0083	2	0.07	7.84
1 lb	2	0.0316	n	0.0316	0.0083	2	0.07	7.84
1 lb	3	-0.0243	n	-0.0243	0.0083	2	0.07	7.84
1 lb	4	0.0004	n	0.0004	0.0083	2	0.07	7.84
1 lb	5	-0.0283	n	-0.0283	0.0083	2	0.07	7.84
0.2 lb		0.0084	n	0.0084	0.0022	2	0.018	7.84
0.2 lb	*	0.0082	n	0.0082	0.0022	2	0.018	7.84
0.1 lb		0.0041	n	0.0041	0.0011	2	0.0091	7.84
0.05 lb		0.00170	n	0.00170	0.00054	2	0.0045	7.84
0.02 lb		0.00029	n	0.00029	0.00022	2	0.0018	7.84
0.02 lb	*	0.00022	n	0.00022	0.00022	2	0.0018	7.84
0.01 lb		0.00039	n	0.00039	0.00018	2	0.0015	7.84
8 oz	11	0.0011	n	0.0011	0.0054	2	0.045	7.84
4 oz	13	0.0003	n	0.0003	0.0028	2	0.023	7.84
2 oz		-0.0025	n	-0.0025	0.0013	2	0.011	7.84
1 oz		0.00188	n	0.00188	0.00064	2	0.0054	7.84

Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly

e-signature is copy only

Joel P. Lavicky Metrologist

8/25/2022

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

Calibration Certificate of Mass

Calibration Date: August 23, 2022		Certificate Number: 2022-114-3	
Submitted By: FSCP Area 20 3721 West Cuming St. Lincoln, NE 68524		Point of Contact: Kurt Wenninghoff Ph. 402-471-3422 email: kurt.wenninghoff@nebraska.gov PO Number: N/a	
Test Item(s): Metric Weight Kit	Artifact(s) Description:	Date Received: 8/19/2022	
Serial Number(s): WM-2-89-4	Condition: Good (some wear)	ID / Asset Number: Area 20	
Material: Stainless Steel	Reference Standards Used:	Class Specification: NIST Class F	
Procedure Used:	Equipment Used:	Manufacture: Troemner	
NSL & /Den Metric Volland-1707	NIST HB 6969, SOP 8 (2019)	Sartorius CC 1201 Mettler XPR 205	
Environmental Cond.	Temp: 21.29 °C	Pressure: 729.6 mmHg	Relative Humidity: 51.07 %
<u>Pertinent Information</u>			
<ul style="list-style-type: none"> The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. RED print indicates an out-of-compliance reading. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and the uncertainty exceed 95% of the maximum permissible error. All of the tolerances and design specifications (except density, hardness and magnetism) were evaluated according to ASTM E617 (2018) and/or NIST HB 105-1 (2019) for the artifacts designated class. All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm³ at 20 °C. It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2020), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales. 			
<u>Traceability Statement</u>			
<p>The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.</p>			
<u>Uncertainty Statement</u>			
<p>The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (<i>k</i>), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the <i>Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)</i>. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.</p>			

Calibration Date: August 23, 2022

Certificate Number: 2022-114-3


Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm ³)
1 kg	1	-0.003	n	-0.003	0.012	2	0.1	7.84
500 g	2	-0.0121	n	-0.0121	0.0083	2	0.07	7.84
200 g	3	-0.0051	n	-0.0051	0.0048	2	0.04	7.84
200 g	4	-0.0127	n	-0.0127	0.0048	2	0.04	7.84
100 g		0.0107	n	0.0107	0.0024	2	0.02	7.84
50 g		0.0074	n	0.0074	0.0012	2	0.01	7.84
20 g		0.00091	n	0.00091	0.00048	2	0.004	7.84
20 g	*	0.00119	n	0.00119	0.00048	2	0.004	7.84
10 g		-0.00107	n	-0.00107	0.00024	2	0.002	7.84

Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



e-signature is copy only

Joel P. Lavicky Metrologist

8/25/2022

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.