

## Calibration Certificate of Mass

**Calibration Date:** May 11, 2020

**Certificate Number:** 2020-053-1

**Submitted By:** FSCP Area 55  
3721 West Cuming St.  
Lincoln, NE 68524

**Point of Contact:** Chris Uglow  
Ph. 402-471-3422  
**email:** [chris.uglow@nebraska.gov](mailto:chris.uglow@nebraska.gov)  
**PO Number:** N/A

<p><b>Test Item(s):</b> (66) Cast Iron Weights <b>Serial Number(s):</b> See Next Page <b>Manufacture:</b> Rice Lake <b>Condition:</b> Good (some wear)</p>	<p><b>Artifact(s) Description:</b></p>	<p><b>Date Received:</b> May 4, 2020 <b>ID / Asset Number:</b> FSCP Area 55 <b>Class Specification:</b> NIST Class F <b>Material:</b> Cast Iron</p>
--	--	---

<p><b>Reference Standards Used:</b> NSL lb standards</p>	<p><b>Procedure Used:</b> NIST HB 6969, SOP 8 (2018) <b>Metrologist:</b> JPL</p>	<p><b>Equipment Used:</b> Mettler XPR32003 Mettler XP 604</p>
--	--	---

**Environmental Cond.**      **Temp:** 20.8 °C    **Pressure:** 736.092 mmHg    **Relative Humidity:** 45.2 %

### Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. All of the tolerances and specifications were evaluated according to ASTM E617 (2018) and/or NIST HB 105-1 (2019).
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

### Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

### Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: May 11, 2020

Certificate Number: 2020-053-1

### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
15 lb	WM-15-19	-0.054	N	-0.054	0.083	2	0.68	7.2
15 lb	WM-15-20	-0.404	N	-0.404	0.083	2	0.68	7.2
25 lb	NE-23	0.09	N	0.09	0.14	2	1.1	7.2
25 lb	NE-25	0.39	N	0.39	0.14	2	1.1	7.2
25 lb	NE-35	0.03	N	0.03	0.14	2	1.1	7.2
25 lb	NE-36	0.61	N	0.61	0.14	2	1.1	7.2
25 lb	NE-37	0.42	N	0.42	0.14	2	1.1	7.2
25 lb	NE-38	0.10	N	0.10	0.14	2	1.1	7.2
25 lb	NE-39	0.78	N	0.78	0.14	2	1.1	7.2
25 lb	NE-40	0.63	N	0.63	0.14	2	1.1	7.2
25 lb	WM-33	-0.22	N	-0.22	0.14	2	1.1	7.2
25 lb	WM-46	-0.79	N	-0.79	0.14	2	1.1	7.2
25 lb	WM-D17	-0.48	N	-0.48	0.14	2	1.1	7.2
25 lb	WM-D18	-0.28	N	-0.28	0.14	2	1.1	7.2
25 lb	WM-D19	-0.05	N	-0.05	0.14	2	1.1	7.2
25 lb	WM-D20	-0.02	N	-0.02	0.14	2	1.1	7.2
25 lb	WM-D27	0.12	N	0.12	0.14	2	1.1	7.2
25 lb	WM-D31	-0.39	N	-0.39	0.14	2	1.1	7.2
25 lb	WM-D34	-0.06	N	-0.06	0.14	2	1.1	7.2
25 lb	WM-D35	-0.50	N	-0.50	0.14	2	1.1	7.2
25 lb	WM-D36	0.18	N	0.18	0.14	2	1.1	7.2
25 lb	WM-D37	0.53	N	0.53	0.14	2	1.1	7.2
25 lb	WM-D37	-0.71	N	-0.71	0.14	2	1.1	7.2
25 lb	WM-D38	-0.11	N	-0.11	0.14	2	1.1	7.2
25 lb	WM-D39	0.26	N	0.26	0.14	2	1.1	7.2
25 lb	WM-D40	-0.19	N	-0.19	0.14	2	1.1	7.2
25 lb	WM-D41	-0.07	N	-0.07	0.14	2	1.1	7.2
25 lb	WM-D42	-0.35	N	-0.35	0.14	2	1.1	7.2
25 lb	WM-D43	0.26	N	0.26	0.14	2	1.1	7.2
25 lb	WM-D45	-0.76	N	-0.76	0.14	2	1.1	7.2
25 lb	WM-D47	-0.74	N	-0.74	0.14	2	1.1	7.2
25 lb	WM-D48	-0.54	N	-0.54	0.14	2	1.1	7.2
25 lb	WM-D49	1.73	Y	-0.21	0.14	2	1.1	7.2
25 lb	WM-D50	-0.12	N	-0.12	0.14	2	1.1	7.2
50 lb	WM-OPI-C61	0.21	N	0.21	0.28	2	2.3	7.2
50 lb	WM-OPI-C64	1.05	N	1.05	0.28	2	2.3	7.2
50 lb	WM-OPI-C65	-1.08	N	-1.08	0.28	2	2.3	7.2
50 lb	WM-OPI-C71	0.08	N	0.08	0.28	2	2.3	7.2
50 lb	WM-OPI-C74	-0.80	N	-0.80	0.28	2	2.3	7.2
50 lb	WM-OPI-C84	1.05	N	1.05	0.28	2	2.3	7.2
1000 lb	B10	-4.6	N	-4.6	5.6	2.009	45	7.2
1000 lb	B12	-5.3	N	-5.3	5.6	2.009	45	7.2
1000 lb	B13	-38.3	Y	-6.2	5.6	2.009	45	7.2
1000 lb	B14	2.7	N	2.7	5.6	2.009	45	7.2
1000 lb	B17	-28.4	N	-28.4	5.6	2.009	45	7.2
1000 lb	B19	11.9	N	11.9	5.6	2.009	45	7.2
1000 lb	B2	11.3	N	11.3	5.6	2.009	45	7.2
1000 lb	B20	-11.3	N	-11.3	5.6	2.009	45	7.2
1000 lb	B21	-11.4	N	-11.4	5.6	2.009	45	7.2
1000 lb	B23	2.8	N	2.8	5.6	2.009	45	7.2

Calibration Date: May 11, 2020

Certificate Number: 2020-053-1


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
1000 lb	B3	2.1	N	2.1	5.6	2.009	45	7.2
1000 lb	B4	-3.4	N	-3.4	5.6	2.009	45	7.2
1000 lb	B5	-7.3	N	-7.3	5.6	2.009	45	7.2
1000 lb	B6	9.4	N	9.4	5.6	2.009	45	7.2
1000 lb	B7	1.0	N	1.0	5.6	2.009	45	7.2
1000 lb	B8	-0.5	N	-0.5	5.6	2.009	45	7.2
1000 lb	B-8	32.7	N	32.7	5.6	2.009	45	7.2
1000 lb	B9	-27.7	Y	22.2	5.6	2.009	45	7.2

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

5/19/2020

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

Calibration Date: 5/4/2020

**Certificate of Calibration  
of Volume Transfer**

Certificate Number: 2020-053-2

**Items Submitted:**

Quantity	Nominal Volume	Manufacturer	Type
3	5 gal	SMI	"Special" J Prover

**Submitted By:** FSCP Area 55  
3721 West Cuming St.  
Lincoln, NE 68524

**POC:** Chris Uglov  
402-471-3422  
[chris.uglov@nebraska.gov](mailto:chris.uglov@nebraska.gov)

**Test Results**

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (/°F)	As Found Volume Delivered @ 60 °F	As left Volume Delivered @ 60 °F	Uncertainty (U)	(k)
5 gal	233	SS	0.0000265	5.0008 gal	5.0008 gal	0.0011 gal	2.03
5 gal	234	SS	0.0000265	4.9986 gal	4.9986 gal	0.0011 gal	2.03
5 gal	235	SS	0.0000265	4.9997 gal	4.9997 gal	0.0011 gal	2.03

*The data in this report only applies to those items specifically listed on this report.*

Volume delivered at 60°F after a 30 second pour and 10 second drain for test measures. For provers a 30 second drain time would apply.

**Conversion Factors:**

1 gal = 231 in<sup>3</sup>  
1 gal = 3.785 412 E-03 m<sup>3</sup>

**Traceability Statement:**

The artifact(s) described in this report have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this report is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this report.

**Uncertainty Statement:**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken.

**Pertinent Information:**

The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. All of the tolerances and specifications were evaluated according to NIST HB 105-3 (2010).

**Condition of Item(s) Submitted for Calibration:**

Good

**Laboratory Reference Standard Used:**

5 gal SP NE 1586

**Treatment of Item(s) before Calibration:**

Tested as Found

**Procedure Used:**

NISTIR 7383, SOP 19 (2016)


**Environmental conditions at time of calibration:**

Temp °C	20.8	Humidity %	51.8
Pressure mmHg	729.49		

**Water temperature at time of calibration:**

52.23 °F

**Date Submitted:** 5/4/2020



Joel P. Lavicky, Metrologist

5/19/2020

Date:

*This document does not represent or imply endorsement by the State of Nebraska, the Nebraska Standards Laboratory or NIST. This document may not be reproduced, except in full, without the written permission of the Nebraska Standards Laboratory*

Calibration Date: 5/5/2020

**Certificate of Calibration  
of Volume Transfer**

Certificate Number: 2020-053-3

**Items Submitted:**

Quantity	Nominal Volume	Manufacturer	Type
2	5 gal	Seraphin	Test Measure 4" Neck

**Submitted By:** FSCP Area 55

3721 West Cuming St.  
Lincoln, NE 68524

**POC:** Chris Uglow

402-471-3422

[chris.uglow@nebraska.gov](mailto:chris.uglow@nebraska.gov)

**Test Results**

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (°F)	As Found Volume Delivered @ 60 °F	As left Volume Delivered @ 60 °F	Uncertainty (U)	(k)
5 gal	4383-5-A	SS	0.0000265	<b>4.9998 gal</b>	<b>4.9998 gal</b>	0.0012 gal	2.08
5 gal	4383-5-B	SS	0.0000265	<b>5.0008 gal</b>	<b>5.0008 gal</b>	0.0012 gal	2.08

*The data in this report only applies to those items specifically listed on this report.*

Volume delivered at 60°F after a 30 second pour and 10 second drain for test measures. For provers a 30 second drain time would apply.

**Conversion Factors:**

1 gal = 231 in<sup>3</sup>

1 gal = 3.785 412 E-03 m<sup>3</sup>

**Traceability Statement:**

The artifact(s) described in this report have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this report is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this report.

**Uncertainty Statement:**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken.

**Pertinent Information:**

The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. All of the tolerances and specifications were evaluated according to NIST HB 105-3 (2010).

**Condition of Item(s) Submitted for Calibration:**

Good

**Laboratory Reference Standard Used:**

5 gal SP NE 1586

**Treatment of Item(s) before Calibration:**

Tested as Found

**Procedure Used:**

NISTIR 7383, SOP 19 (2016)

**Environmental conditions at time of calibration:**


Temp °C	19.4	Humidity %	51.8
Pressure mmHg	731.01		

**Water temperature at time of calibration:**

53.02 °F

**Date Submitted:**

5/5/2020



Joel P. Lavicky, Metrologist

5/19/2020

Date:

*This document does not represent or imply endorsement by the State of Nebraska, the Nebraska Standards Laboratory or NIST. This document may not be reproduced, except in full, without the written permission of the Nebraska Standards Laboratory*



## Calibration Certificate of Mass

<b>Calibration Date:</b> May 7, 2020	<b>Certificate Number:</b> 2020-053-4
--------------------------------------	---------------------------------------

<p><b>Submitted By:</b> FSCP Area 55          3721 West Cuming St.          Lincoln, NE 68524</p>	<p><b>Point of Contact:</b> Chris Uglow          Ph. 402-471-3422  <b>email:</b> <a href="mailto:chris.uglow@nebraska.gov">chris.uglow@nebraska.gov</a>  <b>PO Number:</b> N/A</p>
---	--

<b>Test Item(s):</b> Metric weight kit	<b>Artifact(s) Description:</b>	<b>Date Received:</b> 5/4/2020
<b>Serial Number(s):</b> WM-1G14		<b>ID / Asset Number:</b> FSCP Area 55
<b>Manufacture:</b> Rice Lake		<b>Class Specification:</b> NIST Class F
<b>Condition:</b> Good (some wear)		<b>Material:</b> Stainless Steel

<b>Reference Standards Used:</b>	<b>Procedure Used:</b>	<b>Equipment Used:</b>
OPI & /Den Metric	NIST HB 6969, SOP 8 (2018)	Sartorius CC10000S    Mettler AT 106
	<b>Metrologist:</b>	Sartorius CC 1201    Sartorius CCE6
	JPL	

**Environmental Cond.**    **Temp:** 20.4 °C    **Pressure:** 732.7 mmHg    **Relative Humidity:** 51.2 %

### Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. All of the tolerances and specifications were evaluated according to ASTM E617 (2018) and NIST HB 105-1 (2019).
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

### Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

### Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: May 7, 2020

Certificate Number: 2020-053-4


**Calibration Results**

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
4 kg	6H85	0.091	n	0.091	0.048	2	0.4	7.84
2 kg		0.071	n	0.071	0.024	2	0.2	7.84
1 kg		0.030	n	0.030	0.012	2	0.1	7.84
500 g		0.0207	n	0.0207	0.0083	2	0.07	7.84
200 g		0.0109	n	0.0109	0.0048	2	0.04	7.84
200 g	*	0.0097	n	0.0097	0.0048	2	0.04	7.84
100 g		0.0060	n	0.0060	0.0024	2	0.02	7.84
50 g		0.0026	n	0.0026	0.0012	2	0.01	7.84
20 g		0.00106	n	0.00106	0.00048	2	0.004	7.84
20 g	*	0.00097	n	0.00097	0.00048	2	0.004	7.84
10 g		0.00071	n	0.00071	0.00024	2	0.002	7.84
5 g		0.00042	n	0.00042	0.00018	2	0.0015	7.84
2 g		0.00030	n	0.00030	0.00013	2	0.0011	7.84
2 g	*	0.00024	n	0.00024	0.00013	2	0.0011	7.84
1 g		0.00022	n	0.00022	0.00011	2	0.0009	7.84

**Conversion Factors**

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



**Joel P. Lavicky Metrologist**

**5/27/2020**

**Date of Issue**

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

## Calibration Certificate of Mass

**Calibration Date:** May 8, 2020

**Certificate Number:** 2020-053-5

**Submitted By:** FSCP Area 55  
3721 West Cuming St.  
Lincoln, NE 68524

**Point of Contact:** Chris Uglow  
Ph. 402-471-3422  
**email:** [chris.uglow@nebraska.gov](mailto:chris.uglow@nebraska.gov)  
**PO Number:** N/A

**Test Item(s):** lb weight kit  
**Serial Number(s):** WM-2A86  
**Manufacture:** Rice Lake  
**Condition:** Good (some wear)

**Artifact(s) Description:**

**Date Received:** May 4, 2020  
**ID / Asset Number:** FSCP Area 55  
**Class Specification:** NIST Class F  
**Material:** SS & AL

**Reference Standards Used:**

NSL lb standards

**Procedure Used:**

NIST HB 6969, SOP 8 (2018)

**Metrologist:**

JPL

**Equipment Used:**

Sartorius CC10000S    Mettler AT 106  
Sartorius CC 1201    Sartorius CCE6

**Environmental Cond.**    **Temp:** 19.3 °C    **Pressure:** 737.1 mmHg    **Relative Humidity:** 53.4 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. All of the tolerances and specifications were evaluated according to ASTM E617 (2018) and NIST HB 105-1 (2019).
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.



Calibration Date: May 8, 2020

Certificate Number: 2020-053-5


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
5 lb	1	-0.145	n	-0.145	0.028	2	0.23	7.84
5 lb	2	-0.126	n	-0.126	0.028	2	0.23	7.84
5 lb	3	-0.135	n	-0.135	0.028	2	0.23	7.84
5 lb	4	-0.152	n	-0.152	0.028	2	0.23	7.84
5 lb	5	-0.109	n	-0.109	0.028	2	0.23	7.84
1 lb	1	-0.0228	n	-0.0228	0.0083	2	0.07	7.84
1 lb	2	-0.0090	n	-0.0090	0.0083	2	0.07	7.84
1 lb	3	-0.0283	n	-0.0283	0.0083	2	0.07	7.84
1 lb	4	-0.0025	n	-0.0025	0.0083	2	0.07	7.84
1 lb	5	0.0003	n	0.0003	0.0083	2	0.07	7.84
8 oz		0.0172	n	0.0172	0.0054	2	0.045	7.84
4 oz		0.0064	n	0.0064	0.0028	2	0.023	7.84
2 oz		0.0049	n	0.0049	0.0013	2	0.011	7.84
1 oz		0.00281	n	0.00281	0.00064	2	0.0054	7.84
1/2 oz		0.00040	n	0.00040	0.00034	2	0.0028	7.84
1/4 oz		0.00067	n	0.00067	0.00021	2	0.0017	7.84
1/8 oz		0.00015	n	0.00015	0.00016	2	0.0013	7.84
1/16 oz		0.00016	n	0.00016	0.00014	2	0.0011	7.84
1/32 oz		0.00020	n	0.00020	0.00011	2	0.00087	7.84
1/32 oz	*	0.00049	n	0.00049	0.00011	2	0.00087	7.84

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

5/27/2020

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

## Calibration Certificate of Mass

**Calibration Date:** May 8, 2020

**Certificate Number:** 2020-053-6

**Submitted By:** FSCP Area 55  
3721 West Cuming St.  
Lincoln, NE 68524

**Point of Contact:** Chris Uglow  
Ph. 402-471-3422  
**email:** [chris.uglow@nebraska.gov](mailto:chris.uglow@nebraska.gov)  
**PO Number:** N/A

**Test Item(s):** lb weight kit  
**Serial Number(s):** 13A9  
**Manufacture:** Troemner  
**Condition:** Good (some wear)

**Artifact(s) Description:**

**Date Received:** May 4, 2020  
**ID / Asset Number:** FSCP Area 55  
**Class Specification:** NIST Class F  
**Material:** SS & AL

**Reference Standards Used:**

NSL lb standards

**Procedure Used:**

NIST HB 6969, SOP 8 (2018)

**Metrologist:**

JPL

**Equipment Used:**

Sartorius CC 1201 Sartorius CCE6  
Mettler AT 106

**Environmental Cond.**    **Temp:** 19.3 °C    **Pressure:** 737.1 mmHg    **Relative Humidity:** 53.4 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. All of the tolerances and specifications were evaluated according to ASTM E617 (2018) and NIST HB 105-1 (2019).
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: May 8, 2020

Certificate Number: 2020-053-6


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
2 lb	1	0.019	n	0.019	0.011	2	0.091	7.84
2 lb	2	-0.042	n	-0.042	0.011	2	0.091	7.84
2 lb	3	-0.043	n	-0.043	0.011	2	0.091	7.84
2 lb	4	-0.042	n	-0.042	0.011	2	0.091	7.84
2 lb	5	-0.041	n	-0.041	0.011	2	0.091	7.84
2 lb	6	-0.059	n	-0.059	0.011	2	0.091	7.84
2 lb	7	-0.043	n	-0.043	0.011	2	0.091	7.84
2 lb	8	-0.016	n	-0.016	0.011	2	0.091	7.84
2 lb	9	-0.037	n	-0.037	0.011	2	0.091	7.84
2 lb	10	-0.039	n	-0.039	0.011	2	0.091	7.84
2 lb	11	-0.034	n	-0.034	0.011	2	0.091	7.84
2 lb	12	-0.015	n	-0.015	0.011	2	0.091	7.84
2 lb	13	-0.059	n	-0.059	0.011	2	0.091	7.84
2 lb	14	-0.047	n	-0.047	0.011	2	0.091	7.84
1 lb	15	-0.0174	n	-0.0174	0.0083	2	0.07	7.84
1 lb	16	-0.0195	n	-0.0195	0.0083	2	0.07	7.84
0.3 lb		0.0026	n	0.0026	0.0032	2	0.027	7.84
0.2 lb		-0.0018	n	-0.0018	0.0022	2	0.018	7.84
0.1 lb		-0.0017	n	-0.0017	0.0011	2	0.0091	7.84
0.05 lb		-0.00041	n	-0.00041	0.00054	2	0.0045	7.84
0.03 lb		-0.00126	n	-0.00126	0.00032	2	0.0027	7.84
0.02 lb		0.00042	n	0.00042	0.00022	2	0.0018	7.84
0.01 lb		-0.00024	n	-0.00024	0.00018	2	0.0015	7.84
0.005 lb		-0.00002	n	-0.00002	0.00014	2	0.0012	2.7
0.003 lb		0.00080	n	0.00080	0.00012	2	0.00099	2.7
0.002 lb		0.00062	n	0.00062	0.00011	2	0.00087	2.7
0.001 lb		0.000114	n	0.000114	0.000083	2	0.0007	2.7
0.001 lb		-0.000298	n	-0.000298	0.000083	2	0.0007	2.7
8 oz		-0.0120	n	-0.0120	0.0054	2	0.045	7.84
4 oz		0.0024	n	0.0024	0.0028	2	0.023	7.84
2 oz		0.0016	n	0.0016	0.0013	2	0.011	7.84
1 oz		0.00144	n	0.00144	0.00064	2	0.0054	7.84
1/2 oz		-0.00071	n	-0.00071	0.00034	2	0.0028	7.84
1/4 oz		-0.00026	n	-0.00026	0.00021	2	0.0017	7.84
1/8 oz		-0.00005	n	-0.00005	0.00016	2	0.0013	7.84
1/16 oz		0.00058	n	0.00058	0.00014	2	0.0011	7.84
1/16 oz		0.00033	n	0.00033	0.00014	2	0.0011	7.84

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

5/27/2020

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.